

Serial No. C 118549

Weightron Bilanciai Ltd,
Titan Works
Bridge Way off Broombank Road,
Chesterfield Trading Estate,
Chesterfield S41 9QJ, UK
Tel: 01246 260062
Fax: 01246 260844
www.weightron.com



Weightron Bilanciai Ltd,
Unit 4b, Block 4
Westpark, Chelston
Wellington TA21 9AD, UK
Tel: 01823 662355
Fax: 01823 662017
www.weightron.com



CALIBRATION CERTIFICATE

This Calibration Certificate is issued in accordance with the UKWF Code of Practice for the calibration of Non Automatic Weighing Machines. All weights used are traceable to NPL

Customer

Clews Recycling Ltd
Unit 17, Nutters Lane
Rugby
CV21 1EA
LOCATION
WIR office

Machine Details

Cust ref	W1516078
Make	Bilanciai
Model	DD700
Ser No	2421078
Capacity	50000kg
e/d	12kg
Working cap	44000

Test Weights

Set No.	Cert No.	Date Cal
PB	Tr112424	10/10/25

Cert Interval 12 mths
Next Cert Nov 2025

PP
Serial number

Tolerance +/-	Kg
0-10000	30
10000-40000	40
40000-50000	60

Comments
 $P.H. = 5080$
 $T.W. = 19000$

2	4	6
1	3	5

AS FOUND LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev

DEFINITIVE LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev
400	400	0	0	0
2000	2000	0	400	0
10000	10000	0	2000	0
20000	20000	0	10000	0
25080	25080	0	20000	0
19000	19000	0	19000	0
29000	29000	0	29000	0
39000	39000	0	39000	0
44100	44100	0		

Repeatability

Load	Ind	Ind	Ind

Repeatability

Load	Ind	Ind	Ind
39000	39000	39000	39000

Eccentric Load

Reading	Dev	Reading	Dev
1		2	
3		4	
5		6	
7		8	

Eccentric Load 10000

Reading	Dev	Reading	Dev		
1	10000	0	2	10000	0
3	10000	0	4	10000	0
5	10000	0	6	10000	0
7			8		

Signature *[Signature]*

Name *D. Parkes*

Date *14/11/2025*